## Zibo Eastmed Healthcare Products Co., Ltd.

Address: No.118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

## ANALYSIS CERTIFICATE

Item Name	Disposable 5ml / cc Luer lock syringes		Needle Size		22GX1 1/2"			
Lot No.	20221116			Lot Qty		271200pcs		
MFG DATE	20221116			Inspection Item		Biological/Chemical/Physical		
EXP DATE	DATE 20271115		Inspection Date		ate	2022.12.10		
Standards	ISO8537/YY0497			Report Date		2022.12.15		
Date of elaboration:	202	022.12.15 Expiration date			ite	2027.11.15		
Item			Requirements		R	esult	Conclusion	
Biological								
Sterility		Shall be sterility,no bacteria				Sterile	Pass	
Pyrogenicity		Shall be non-pyrogenicity			Non	-pyrogenicity	Pass	
Chemical								
Limits for acidity or alkalinity		the difference of PH value should be no more than 1				0.18	Pass	
Limits for extractable metals		the color of the inspection solution should not be deeper than the color of the blank reference solution			per	Conforms	Pass	
Physical								
Tolerance on graduate capacity		Shall be conform the requirement			(	Conforms	Pass	
Freedom from		Shall be clean and free from extraneous matter when viewed			/ed	- ·	Pass	
extraneous matter		by normal or corrected vision without magnification				Conforms		
Graduated lines		Shall be conform the requirement			(	Conforms	Pass	
Piston/plunger assembly		Shall be conform the requirement			(	Conforms	Pass	
Lubricant		Shall not be visible droplets or particles under normal or corrected-to-normal			or (	Conforms	Pass	
Dead space		Shall be conform the requirement			0.00	25~0.0039ml	Pass	
Needle point		Shall appear sharp,no burrs and hooks				Conforms	Pass	
Bond between hub and and needle tube		Shall not be broken by the 22 N force			(	Conforms	Pass	
Needle Point pund pressure	Point puncture Not more than <u>0.7</u> N		(	Conforms	Pass			
Needle tube size		Shall be conform	Il be conform Length: 11~14m		11.9	11.93~13.09mm Pass		
		requirement External diameter: 0.335~0.345mm		0.33	37~0.340mm	Pass		
Needle tube bore		Shall be clean;no dirt through the tube bore				Conforms Pass		
Connect effect between		Shall be straightness,no deflection clearly				Conforms Pass HEALTHCARE PROOF		
needle hub and tube					HEA			
Conclusion PASS								