Zibo Eastmed Healthcare Products Co., Ltd.

Address: No.118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

ANALYSIS CERTIFICATE

Item Name	Disposable 20ml / cc Luer Lock syringes		Needle Size		21GX 1 ½ "		
Lot No.	20221116			Lot Qty		225000pcs	
MFG DATE	20221116			Inspection Item		Biological/Chemical/Physical	
EXP DATE	20271115			Inspection Date		2022.12.10	
Standards	ISO8537/YY0497			Report Date		2022.12.15	
Date of elaboration:	202	2.12.15		Expiration da	te	2027.11.15	
Item		Requirements		R	Result Conclusion		
Biological							
Sterility		Shall be sterility,no bacteria				Sterile Pass	
Pyrogenicity		Shall be non-pyrogenicity			Non	lon-pyrogenicity Pass	
Chemical							
Limits for acidity or alkalinity		the difference of PH value should be no more than 1				0.18	Pass
Limits for extractable		the color of the inspection solution should not be deeper			per (Conforms Pass	
metals than the color of the blank reference solution							
Physical							
Tolerance on graduate capacity		Shall be conform the requirement			(Conforms	Pass
Freedom from Sha		Shall be clean and free from extraneous matter when viewed			ed	Conforms	Pass
extraneous matter		by normal or corrected vision without magnification			`	50111011113	
Graduated lines		Shall be conform the requirement			(Conforms Pass	
Piston/plunger assembly		Shall be conform the requirement			(Conforms Pass	
Lubricant		Shall not be visible droplets or particles under normal or corrected-to-normal			or	Conforms	Pass
Dead space		Shall be conform the requirement			0.00	25~0.0039ml	Pass
Needle point	edle point Shall appear sharp,no burrs and hooks		,no burrs and hooks		(Conforms	Pass
Bond between hub and and needle tube		Shall not be broker	all not be broken by the22 N force			Conforms F	
Needle Point puncture pressure		Not more than 0	Not more than <u>0.7</u> N			Conforms	Pass
Needle tube size		Shall be conform	Length: 11~14m		11.9	3~13.09mm	Pass
INGCUIC LUDE SIZE		requirement	External diameter: 0.3	35~0.345mm	0.33	37~0.340mm	Pass
Needle tube bore		Shall be clean;no dirt through the tube bore			(Conforms Pass	
Connect effect between needle hub and tube		Shall be straightness,no deflection clearly			(Conforms	Pass
Conclusion						HEALTHOA!	E PRODU
Examiner: Liu Fang Checker: Li Gang Checker: Li Gang							