## Zibo Eastmed Healthcare Products Co., Ltd.

Address: No.118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

## ANALYSIS CERTIFICATE

Lot No. MFG DATE	Disposable 10ml / 20210326	cc Luer Lock syringes	Needle Size		21GX 1 ½	,,	
	20210326		Needle Size		21GX 1 ½ "		
MFG DATE			Lot Qty		400000pcs		
	20210326		Inspection Item		Biological/Chemical/Physical		
EXP DATE	20260325		Inspection Date		2021.04.10		
Standards	ISO8537/YY0497		Report Date		2021.04.15		
Date of elaboration:	2021.03.26	1.03.26 Expiration date			2026.03.25		
Item		Requirements			Result Conclusion		
Biological		rtoquiromonto				- Contraction	
Sterility Shall be sterility,no bacteria					Sterile Pass		
Pyrogenicity		Shall be non-pyrogenicity			-pyrogenicity	Pass	
Chemical		pyrogemeny			p)3		
Limits for acidity alkalinity	or the difference	the difference of PH value should be no more than 1			0.18	Pass	
Limits for extracta		the color of the inspection solution should not be deeper			Conforms	Pass	
Physical	than the colo	of the blank reference solution	JII				
Tolerance on gradu	Shall be conf	Shall be conform the requirement			Conforms	Pass	
		Shall be clean and free from extraneous matter when viewed by normal or corrected vision without magnification			Conforms	Pass	
Graduated lines		Shall be conform the requirement			Conforms	Pass	
Piston/plunger assemb		Shall be conform the requirement			Conforms	Pass	
Lubricant	Shall not be	Shall not be visible droplets or particles under normal or corrected-to-normal			Conforms	Pass	
Dead space	Shall be conf	Shall be conform the requirement			25~0.0039ml	Pass	
Needle point	_	Shall appear sharp,no burrs and hooks			Conforms	Pass	
Bond between hub a	and	Shall not be broken by the N force			Conforms	Pass	
Needle Point punct	Not more than <u>0.7</u> N			Conforms	Pass		
	Shall be conf	form Length: 11~14m		11.9	93~13.09mm	Pass	
Needle tube size	requirement	External diameter: 0.3	335~0.345mm	0.33	37~0.340mm	Pass	
Needle tube bore	Shall be clea	Shall be clean;no dirt through the tube bore			Conforms Pass		
Connect effect between	Shall be strai	Shall be straightness,no deflection clearly			Conforms	Pass	
Conclusion			PASS				