Zibo Eastmed Healthcare Products Co., Ltd.

Address: No.118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

ANALYSIS CERTIFICATE

	1					Walleton and the same of the s	Manual 20	
Item Name	m Name Disposable 3ml / cc Luer Loc		r Lock syringes	Needle Size		23GX 1 1/4 "		
Lot No.	202	10826		Lot Qty		423000pcs	423000pcs	
MFG DATE	DATE 20210826			Inspection Item		Biological/Chemical/Physical		
XP DATE 20260825		9	Inspection Date		2021.09.10			
tandards ISO8537/YY0497		*	Report Date		2021.09.15			
Date of elaboration:	202	1.8.26	Expiration date		2026.03.25			
Item		Requirements			Result	Conclusion		
Biological			- toquiromonio					
Sterility		Shall be sterility,no bacteria				Sterile	Pass	
Pyrogenicity		Shall be non-pyrogenicity			N	Non-pyrogenicity	Pass	
Chemical							L	
Limits for acidity or alkalinity		the difference of PH value should be no more than 1				0.18	Pass	
Limits for extractable metals		the color of the inspection solution should not be deeper than the color of the blank reference solution			eper	Conforms	Pass	
Physical						A SUMMER SECTION OF A VALUE OF THE SECTION OF THE S		
Tolerance on graduate capacity		Shall be conform the requirement				Conforms	Pass	
Freedom from extraneous matter		Shall be clean and free from extraneous matter when viewed by normal or corrected vision without magnification			wed	Conforms	Pass	
Graduated lines		Shall be conform the requirement				Conforms	Pass	
Piston/plunger assembly		Shall be conform the requirement				Conforms	Pass	
Lubricant		Shall not be visible droplets or particles under normal or corrected-to-normal			l or	Conforms	Pass	
Dead space		Shall be conform the requirement			0	.0025~0.0039ml	Pass	
Needle point		Shall appear sharp,no burrs and hooks			777	L Spnforms	Pass	
Bond between hub and and needle tube		Shall not be broken by the 22 N force			3	Conforms	Pass	
Needle Point pun-	Not more than <u>0.7</u> N			Wall House of the last	Conforms	Pass		
Needle tube size		Shall be conform Length: 11~14m requirement External diameter: 0.335~0.345mm				11.93~13.09mm	Pass	
				(0.337~0.340mm	Pass		
Needle tube bore		Shall be clean;no dirt through the tube bore				Conforms	Pass	
Connect effect between needle hub and tube		Shall be straightness,no deflection clearly				Conforms	Pass	
Conclusion	PASS							