

ZIBO EASTMED HEALTHCARE PRODUCTS CO.,LTD

Latex Exam Gloves

Certificate of Analysis

Powdered

Product Name		Non Sterile Powder Latex Exam Gloves	Product Standard		EN455-1, -2,-3
Lot No.		200611	Type, Size		S
Mfg date		06/2020	Exp date		05/2025
Lot Quantity		500000 Pcs	Sample Quantity		315 PCS
Sampling		10-06-20~10-06-20	Test Date		10-06-20~17-06-20
Sampler		Wang Huihui	Identifier		Xiao Gang
Test Item		Index	Test Level	Test Result	Result
Before Aging	Tensile Strength	≥9 N	S-2 AQL 1.5	9.7~11.6, 0/13	Pass
Aged	Tensile Strength	≥6N	S-2 AQL 1.5	7.6~9.2, 0/13	Pass
S	Length	≥230 mm	S-2 AQL 1.5	240~243, 0/13	Pass
	Width	85±5mm	S-2 AQL 1.5	84~90, 0/13	Pass
	Thickness	≥0.08mm	S-2 AQL 1.5	0.09~0.1,0/13	Pass
Leachable Protein		200µg/dm ²	3 pcs	75.0 µg/dm ²	Pass
Freedom from holes			G-1 AQL1.5	3/315	Pass
Labeling			S-2 AQL 1.5	0/13	Pass
Biological Test			Non-Irritation to Rabbit's Skin and Non-sensitization to Guinea Pig's Skin.		
Determinant Compliance			This lot conforms to the requirements established by the specifications of EN455-1, EN455-2 , EN455-3		

Approved By Xiao Gang

