Zibo Eastmed Healthcare Products Co., Ltd.

Address: No. 118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

ANALYSIS CERTIFICATE

				T		1		
Item Name	Dis	posable 10ml / cc Lu	Needle Size		21GX 1 ½	21GX 1 ½"		
Lot No.	202	00326		Lot Qty		278400pcs		
MFG DATE	202	00326		Inspection Item		Biological/Chemical/Physical		
EXP DATE	202	50325		Inspection Date		2020.5.20		
Standards	ISO	08537/YY0497		Report Date		2020.5.20	2020.5.20	
Date of elaboration:	202	20.3.26 Expiration da			te	2025.3.25		
Item		Requirements				_⊦ Result	Conclusion	
Biological								
Sterility Shall be sterility			no bacteria			Sterile	Pass	
Pyrogenicity		Shall be non-pyrogenicity			No	n-pyrogenicity	Pass	
Chemical							I	
Limits for acidity or alkalinity		the difference of PH value should be no more than 1				0.18	Pass	
Limits for extractable metals		the color of the inspection solution should not be deeper than the color of the blank reference solution			per	Conforms	Pass	
Physical								
Tolerance on graduate capacity		Shall be conform the requirement				Conforms	Pass	
Freedom from extraneous matter		Shall be clean and free from extraneous matter when viewed by normal or corrected vision without magnification			/ed	Conforms	Pass	
Graduated lines		Shall be conform the requirement				Conforms	Pass	
Piston/plunger assembly		Shall be conform the requirement				Conforms	Pass	
Lubricant		Shall not be visible droplets or particles under normal or corrected-to-normal			or	Conforms	Pass	
Dead space		Shall be conform the requirement			0.00	025~0.0039ml	Pass	
Needle point S		Shall appear sharp,no burrs and hooks				Conforms	Pass	
Bond between hub and and needle tube		Shall not be broken by the22 N force				Conforms	Pass	
Needle Point pund	ture	Not more than <u>0.7</u> N				Conforms	Pass	
Needle tube size		Shall be conform	Length: 11~14m		11.	93~13.09mm	Pass	
indecile tube Size		requirement	External diameter: 0.3	35~0.345mm	0.3	37~0.340mm	Pass	
Needle tube bore		Shall be clean;no dirt through the tube bore				Conforms	Pass	
Connect effect betw needle hub and tube	Connect effect between eedle hub and tube Shall be straightness,no deflection clearly					Conforms	Pass	
Conclusion		PASS						
					Caul I			

Authorized Signature(s)