## Zibo Eastmed Healthcare Products Co., Ltd.

Address: No.118, Huaguang Road, Zhangdian District, Zibo City, Shandong Province, China

## ANALYSIS CERTIFICATE

Item Name	Disposable 1ml / cc Lue	Disposable 1ml / cc Luer Slip syringes		27GX 1/2"	27GX 1/2"	
Lot No.	20220426		Lot Qty	144000pcs	144000pcs	
MFG DATE	20220426		Inspection Item	Biological/C	Biological/Chemical/Physical	
EXP DATE	P DATE 20270425		Inspection Date	2022.07.10	2022.07.10	
Standards	ISO8537/YY0497		Report Date	2022.07.15	2022.07.15	
Date of elaboration:	2022.07.15	2.07.15 Expiration		te 2027.04.25		
Item		Requirements		Result	Conclusion	
Biological						
Sterility Shall be sterility,no bacteria			Sterile	Pass		
Pyrogenicity		Shall be non-pyrogenicity		Non-pyrogenicity	Pass	
Chemical						
Limits for acidity alkalinity	or the difference of P	the difference of PH value should be no more than 1			Pass	
Limits for extract		the color of the inspection solution should not be deeper than the color of the blank reference solution			Pass	
Physical						
Tolerance on grad	Shall be conform t	Shall be conform the requirement			Pass	
		Shall be clean and free from extraneous matter when viewed by normal or corrected vision without magnification			Pass	
Graduated lines		Shall be conform the requirement			Pass	
Piston/plunger assem	nbly Shall be conform t	Shall be conform the requirement			Pass	
Lubricant		Shall not be visible droplets or particles under normal or corrected-to-normal			Pass	
Dead space	Shall be conform t	Shall be conform the requirement			Pass	
Needle point	Shall appear sharp	Shall appear sharp,no burrs and hooks		Conforms	Pass	
Bond between hub and needle tube	and	Shall not be broken by the22 N force		Conforms	Pass	
Needle Point pund	edle Point puncture Not more than 0.7 N			Conforms	Pass	
Needle tube size	Shall be conform requirement	Length: 11~14m	* *	11.93~13.09mm	Pass	
		External diameter:	0.335~0.345mm	0.337~0.340mm	Pass	
Needle tube bore	Needle tube bore Shall be clean;no dir		irt through the tube bore		Pass	
Connect effect betweeneedle hub and tube	Shall be straightness, no deflection clearly		Conforms	Pass		
Conclusion			PASS	2		
Examiner:	Liu Fang	OV	ecker: Li Gang			

Examiner:

Liu Fang

Checker:

Li Gang