

# ZIBO EASTMED HEALTHCARE PRODUCTS CO LTD

## Test Report

Name of Tested Products	DISPOSABLE SYRINGE	Specification and Type	5ml 22G x 1 1/2''		
Lot No.	20240102	Exp Date	01-01-2029		
Fab Date.	2024-01-02	Quantity	204000pcs		
Testing Accordance	GB15810-2001		Insect every batch		
Date of Testing	2024.01.06	Date of Report	2024.01.09		
	Testing Items	Technical Requirements	Testing Conclusion	Conclusion	
Category A	Asepsis	Should be asepsis	Asepsis	Qualified	
	Heat source	Should be no heat source	No heat source	Qualified	
	EO remain volume	The remain volume should $\leq$ 10ug/g	0.77ug/g	Qualified	
	Testing Items	Technical Requirements	Sampling Scheme	Number that not up to standard	Judgement
			n		
Category B	Connect of taper	Accord with the regulation of GB1962.1 or GB1962.2	20	0	Qualified
	Syringe Column Sealing	300Kpa Positive Pressure :No Leakaging; 88Kpa Negative Pressure No Leakaging			
	The permit difference of capacity	$\leq \pm 4\%$			
Category C	Appearance	Should be clean, transparent , no bur, no semi-finished edge, no lubricating oil assemble.	20	0	Qualified
	Gauge	The length of gauge $\geq$ 36 mm			
	Indicating Ruler,0 level line	Within 1/4 range			
	The print of gauge	The position of gauge should right, the print complete.			
	The rubber plug	Should no impurities			
	The slide function	Should have the good slide function			
	Remain volume	Remain volume $\leq$ 0.075ml			
	The primary pack	The symbol of primary pack should complete.			
Conclusion	Qualified		Inspect stamp:		

Inspector: Jiao jing      Checking: Xu huijing      Examination and Approval: Jiang chungin

For and on behalf of  
 ZIBO EASTMED HEALTHCARE PRODUCTS CO., LIMITED  
  
 .....  
 Authorized Signature(s)